

High Plains Radiological Assn

1901 MediPark Dr. #2050, Amarillo, Texas, 79106

PROTOCOL	CT MAXILLOFACIAL
SERIES #	1
PARAMETERS	
Slice thickness(mm)	1.25
Index (interval)	1.25
FOV	15-18CM
Algorithm	STD/BONE
CONTRAST	
Oral	NO
IV	ONLY FOR TUMOR
Amount	100CC
Bolus	
Delay/trigger	
SCANNING	
Start level	TOP OF FRONTAL SINUS
Stop level	THROUGH MANDIBLE
Inspiration	NO
Expiration	NO
WINDOW RECONS	
Soft tissue	YES
Lung	NO
Bone	NO
Reformations	SAG/COR MPR
NOTES	